

# Physicians say non-contact infrared thermometers fall short as COVID-19 screeners

December 15 2020

---



Physicians at Johns Hopkins Medicine and the University of Maryland Medical School say a non-contact infrared thermometer, such as the one being used here to check a traveler for fever at the airport, is a poor means of screening for COVID-19 infection. Credit: Public domain

While a fever is one of the most common symptoms for people who get

sick with COVID-19, taking one's temperature is a poor means of screening who is infected with the SARS-CoV-2 virus that causes the disease, and more importantly, who might be contagious. That's the conclusion of a perspective editorial by researchers at Johns Hopkins Medicine and the University of Maryland School of Medicine that describes why temperature screening—primarily done with a non-contact infrared thermometer (NCIT)—doesn't work as an effective strategy for stemming the spread of COVID-19.

The editorial was published Dec. 14, 2020, in *Open Forum Infectious Diseases*, the online journal of the Infectious Diseases Society of America. The authors are William Wright, D.O., M.P.H., assistant professor of medicine at the Johns Hopkins University School of Medicine, and Philip Mackowiak, M.D., M.B.A., emeritus professor of medicine at the University of Maryland School of Medicine.

In March 2020, the U.S. Department of Health and Human Services and the U.S. Centers for Disease Control and Prevention released guidelines for Americans to determine if they needed to seek medical attention for symptoms suggestive of infection with SARS-CoV-2, with [temperature](#) screening playing an integral role. According to the guidelines, fever is defined as a temperature—taken with an NCIT near the forehead—of greater than or equal to 100.4 degrees Fahrenheit (38.0 degrees Celsius) for non-health care settings and greater than or equal to 100.0 degrees Fahrenheit (37.8 degrees Celsius) for health care ones. This is the first aspect of COVID-19 screening by temperature that Wright and Mackowiak question in their editorial.

"Readings obtained with NCITs are influenced by numerous human, environmental and equipment variables, all of which can affect their accuracy, reproducibility and relationship with the measure closest to what could be called the 'body temperature'—the core temperature, or the temperature of blood in the pulmonary vein," says Wright.

"However, the only way to reliably take the core temperature requires catheterization of the pulmonary artery, which is neither safe nor practical as a screening test."

In their editorial, Wright and Mackowiak provide statistics to show that NCIT fails as a screening test for SARS-CoV-2 infection.

"As of Feb. 23, 2020, more than 46,000 travelers were screened with NCITs at U.S. airports, and only one person was identified as having SARS-CoV-2," says Wright. "In a second example, CDC staff and U.S. customs officials screened approximately 268,000 travelers through April 21, 2020, finding only 14 people with the virus."

From a November 2020 CDC report, Wright and Mackowiak provide further support for their concern about temperature screenings for COVID-19. The report, they say, states that among approximately 766,000 travelers screened during the period Jan. 17 to Sept. 13, 2020, only one person per 85,000—or about 0.001%—later tested positive for SARS-CoV-2. Additionally, only 47 out of 278 people (17%) in that group with symptoms similar to SARS-CoV-2 had a measured temperature meeting the CDC criteria for fever.

Another problem with NCITs, Wright says, is that they may give misleading readings throughout the course of a fever that make it difficult to determine when someone is actually feverish or not.

"During the period when a fever is rising, a rise in core temperature occurs that causes blood vessels near the skin's surface to constrict and reduce the amount of heat they release," Wright explains. "And during a fever drop, the opposite happens. So, basing a fever detection on NCIT measurements that measure heat radiating from the forehead may be totally off the mark."

Wright and Mackowiak conclude their editorial by saying that these and other factors affecting thermal screening with NCITs must be addressed to develop better programs for distinguishing people infected with SARS-CoV-2 from those who are not.

Among the strategies for improvement that they suggest are: (1) lowering the cutoff temperature used to identify symptomatic infected people, especially when [screening](#) those who are elderly or immunocompromised, (2) group testing to enable real-time surveillance and monitoring of the virus in a more manageable situation, (3) "smart" thermometers—wearable thermometers paired with GPS devices such as smartphones, and (4) monitoring sewage sludge for SARS-CoV-2.

**More information:** William F Wright et al, Why Temperature Screening for COVID-19 with Non-Contact Infrared Thermometers Doesn't Work, *Open Forum Infectious Diseases* (2020). [DOI: 10.1093/ofid/ofaa603](https://doi.org/10.1093/ofid/ofaa603)

Provided by Johns Hopkins University School of Medicine

Citation: Physicians say non-contact infrared thermometers fall short as COVID-19 screeners (2020, December 15) retrieved 30 June 2024 from <https://medicalxpress.com/news/2020-12-physicians-non-contact-infrared-thermometers-fall.html>

<p>This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.</p>
--