

Clinicians may soon be able to tell COVID-19 from seasonal flu with fast turnaround

January 10 2022



(Top) Rendering of the ball-based valve mechanism that is similar to a ballpoint pen, in which ink is dispensed onto paper when the metal ball at the tip is pressed while writing. The valve is closed when the ball functions as a plug for the well at the top. The valve is open by the pin in the unit at the bottom that lifts the ball up when the pin is aligned with the ball after sliding, releasing the reagent. (Middle) A sample is introduced into the bottom well, followed by sliding the bottom well to have a sequential delivery of 4 reagents for virus lysis and RNA isolation/purification. (Bottom) Colorimetric detection aided by a UV flashlight and the overall time from sample to answer in 50 min. Credit: University of Florida



For those who have made a trip to a doctor or clinic during the oncoming flu season, the most nagging concern is whether their symptoms are indicative of the garden-variety bug that is making the rounds in the workplace or a potential COVID-19 infection. As researchers scramble to treat this virus—and its inevitable variants and mutations—one of the biggest challenges has been quickly and accurately diagnosing the source of the illness so that the proper treatment protocols are administered, and in time to decrease the odds of potentially dangerous complications.

Led by Dr. Z. Hugh Fan, Ph.D., professor at the Herbert Wertheim College of Engineering's Department of Mechanical and Aerospace Engineering, and Dr. John Lednicky, Research Professor at the College of Public Health and Health Professions' Department of Environmental and Global Health, an interdisciplinary team at the University of Florida has developed a game-changing <u>diagnostic test</u> for SARS-CoV-2 that is fast, reliable, low-cost and capable of differentiating between COVID-19 and influenza.

Dr. Fan and his team demonstrated the ability to detect two viruses in 50 minutes without using bulky or sophisticated laboratory equipment or power supply, and with <u>detection sensitivity</u> that is on par with that of the gold-standard RT-PCR (reverse transcription polymerase chain reaction) assay, which is the test recommended by the World Health Organization and the Centers for Disease Control and Prevention. The effort is supported by the National Institutes of Health.

"In patient-care terms, this innovative sample-to-answer platform is an important upgrade from the existing RT-PCR standard on several levels," Dr. Fan said. "With regard to diagnosis, the turnaround time is less than an hour at the testing spot compared to one to two days from an offsite testing lab."





Exploded view of the device consisting of 3 components, including a buffer unit at the top, a mixing unit in the middle, and the detection unit at the bottom. The buffer unit contains 4 wells in each side for the storage of a lysis buffer, a binding buffer, and 2 washing buffers for sample preparation. In each well of the buffer unit one stainless steel ball is placed at the bottom that functions as the valve. The mixing unit has one well in each side and it slides under the buffer unit through the sliding tracks on both sides and in the middle. The detection unit contains a paper pad that is laminated between two thermoplastic films and attached to the well layer using a double-sided adhesive tape. Credit: University of Florida

The UF-engineered device was recently published for peer review in *ACS Sensors*, the flagship journal of the American Chemical Society, focusing on sensor science.

RT-LAMP (loop-mediated isothermal amplification) is an assay used for viral RNA detection. The methodology for RT-LAMP was based on the mechanism behind auto-cycling strand displacement DNA synthesis. A polymerase carries out the reaction, and the polymerase has high strand displacement activity. There are also two pairs of primers used, one pair of inner and another of outer primers. These primers are specially designed for the reaction. RT-LAMP achieves high specificity due to the target sequences. Unlike other technologies, RT-LAMP recognizes the target sequence using six independent sequences at the start, giving it



high specificity. Primer recognition of the target genome leads to a strong colorimetric reaction, which allows for detection without highly specialized or costly instrumentation.

"RT-LAMP technology has numerous advantages over standard RT-PCR," Dr. Lednicky said. "The cost of reagents is low compared to an equivalent RT-PCR reaction, results are obtained more quickly, and equipment needs for RT-LAMP are low, which is a major advantage for users in resource-strapped countries or for use in doctor's offices and clinics. Also, packaged RT-LAMP reactions in a kit format take up little storage space and can be freeze-dried, alleviating the need for freezer storage."



Photograph of an assembled device, in which the mixing unit slides to the second well from the left, lifting up those balls there. (inset) Top view picture of a detection unit, showing a well for RNA isolation and RT-LAMP. Credit: University of Florida

The specimen—swab, urine, saliva, blood, depending on the test—is taken, transferring a determined amount into the first chamber of the sample work-up device. Sequentially, the specimen is subjected to



specific fluids that extract DNA or RNA from the pathogen, purifying the extracted material and collecting it on a detachable filter unit. RT-LAMP chemicals are added into the detachable unit, the chamber with the RT-LAMP fluids is sealed with tape and the final unit is detached and placed into a microchip-controlled "coffee mug," which is preheated to a specific temperature. After a selected incubation period, the results are determined, evidenced by a color change that is visually detected or that can also be detected/measured through its level of fluorescence.

Using disposable-format single-use units, the device can be used by nurses, doctors, and others without the need for advanced training. The results, positive or negative, can be determined visually or with the aid of simple technology like a flashlight. The results can be recorded by using an iPhone or similar camera-equipped mobile device to capture the image, where it is sent to a remote site for confirmation of the data.

"Bottom line, the device allows for quicker analysis of genetic material than traditional RT-PCR and has been successfully used in the detection of the SARS-CoV-2 virus," Dr. Lednicky said. "It's simple, cheap, and does not require advanced technical skills or expensive equipment. Most importantly, it's very sensitive, yielding detection sensitivities, in some cases, 10 times more sensitive than standard RT-PCR assays."





Photographs of commercially available, battery-operated coffee mug used as a water bath for isothermal amplification. (left) Coffee mug is open, containing detection units in a water bath. A picture of the detection unit is shown in the inset above the mug. (right) Coffee mug is closed during RT-LAMP amplification at 62.5 degrees C. Credit: University of Florida

The simultaneous test of both SARS-CoV-2 and influenza viruses isn't presently available at point-of-care (POC), but its variations are available in test labs, and would likely be considered as a "CLIA-waived" device, which includes home tests approved by the FDA. "The device could also be used for <u>environmental studies</u> (SARS-CoV-2 in air, sewage or other matrices), and can be affixed/mated with other detection devices such as air samplers," Dr. Fan said.

The development of this LAMP-based device comes on the heels of progressive advancements by UF researchers in the early detection and identification of dangerous pathogens. Three years ago, with funding from the State of Florida, Dr. Fan and Dr. Lednicky led a research team that developed a rapid, cost-effective miniaturized device for the detection of the mosquito-borne Zika virus, which presents no symptoms in up to 80 percent of those infected. The point-of-care testing platform can be used for screening asymptomatic patients in the field, mitigating



misdiagnosis and helping to prevent further transmission.

Last year, engineers in the Department of Environmental Engineering Sciences at the Engineering School of Sustainable Infrastructure & Environment (ESSIE) led by Dr. Chang-Yu Wu, developed a novel virus aerosol sampler that efficiently collects airborne viruses (including COVID-19)—information from which enables healthcare providers to pinpoint risk areas and identify measures for improvement. The integration of the new sampler with the RT-LAMP device could potentially reach beyond the testing areas into general public areas (e.g. airports, subways, government buildings, shopping centers) providing early warning for airborne pathogenic viruses, helping to stem their spread.

More information: Carlos Manzanas et al, A Valve-Enabled Sample Preparation Device with Isothermal Amplification for Multiplexed Virus Detection at the Point-of-Care, *ACS Sensors* (2021). <u>DOI:</u> <u>10.1021/acssensors.1c01718</u>

Provided by University of Florida

Citation: Clinicians may soon be able to tell COVID-19 from seasonal flu with fast turnaround (2022, January 10) retrieved 28 June 2024 from <u>https://medicalxpress.com/news/2022-01-clinicians-covid-seasonal-flu-fast.html</u>

This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.