

## Researchers develop ultra-sensitive photoacoustic microscopy for wide biomedical application potential

August 31 2023



(a) SLD-PAM system and (b) sensitivity improvement from the probe and filter. Credit: *Advanced Science* (2023). DOI: 10.1002/advs.202302486

Optical-resolution photoacoustic microscopy is an up-and-coming biomedical imaging technique for studying a broad range of diseases, such as cancer, diabetes and stroke. But its insufficient sensitivity has been a longstanding obstacle for its wider application.



Recently, a research team from City University of Hong Kong (CityU) developed a multi-spectral, super-low-dose photoacoustic <u>microscopy</u> system with a significant improvement in the system <u>sensitivity</u> limit, enabling new biomedical applications and clinical translation in the future. The findings were published in the journal *Advanced Science* under the title "Super-Low-Dose Functional and Molecular Photoacoustic Microscopy."

Photoacoustic microscopy is a biomedical imaging technique that combines ultrasound detection and <u>laser</u>-induced photoacoustic signals to create detailed images of biological tissue. When biological tissue is irradiated with a pulsed laser, it generates ultrasonic waves, which are then detected and converted into electric signals for imaging. This technique can achieve up to capillary-level or sub-cellular resolution at greater depths than traditional optical microscopy methods. However, insufficient sensitivity has hindered the technology's wider application.

"High sensitivity is important for high-quality imaging. And it helps detect chromophores (molecules that confer color on materials by absorbing particular wavelengths of visible light) that do not strongly absorb light. It also helps lessen photobleaching and phototoxicity, reduce perturbation to the biological tissues of delicate organs, and broaden the choices of low-cost, low-power lasers in a wide spectrum," explained Professor Wang Lidai, Associate Professor in the Department of Biomedical Engineering at CityU.

For instance, in an ophthalmic examination, a low-power laser is preferred for more safety and comfort. Long-term monitoring of pharmacokinetics or blood flow requires low-dose imaging to alleviate perturbation to tissue functions, he added.





Image comparison of the in vivo results of (a) Traditional-PAM, (b) SLD-PAM at super-low pulse energy with a green-light source, and (c) oxygen saturation image SLD-PAM acquired via dual-wavelength spectrum unmixing. Credit: *Advanced Science* (2023). DOI: 10.1002/advs.202302486

To overcome the sensitivity challenge, Professor Wang and his research team recently developed a multi-spectral, super-low-dose photoacoustic microscopy (SLD-PAM) system that breaks through the sensitivity limit of traditional photoacoustic microscopy, significantly improving sensitivity by about 33 times.

They achieved the breakthrough by combining improvement in the photoacoustic sensor design and innovation of a 4D spectral-spatial filter algorithm for computation. They improved the sensor design by using a lab-customized high-numerical-aperture acoustic lens, optimizing the optical and acoustic beam combiner, and improving the optical and acoustic alignment. The SLD-PAM also utilizes a low-cost multiwavelength pulsed laser, providing 11 wavelengths, ranging from green to red light. The laser operates at a repetition frequency up to megahertz,



and the spectral switching time is in sub-microseconds.

To demonstrate the significance and novelty of SLD-PAM, the team tested it thoroughly via in vivo animal imaging at super-low pulse energy with green-light and red-light sources, resulting in remarkable findings.

First, SLD-PAM enabled high-quality in vivo anatomical and functional imaging. The super-low laser power and high sensitivity significantly reduced perturbations in eye and brain imaging, paving an avenue for clinical translation. Second, without compromising <u>image quality</u>, SLD-PAM reduced photobleaching by about 85%, using lower laser power, and enabled the use of a much broader range of molecular and nanoprobes. In addition, the system cost is significantly lower, making it more affordable for research laboratories and clinics.



Image comparison of the in vivo results of (a) Low-dose PAM and (b) SLD-PAM, both at super-low pulse energy with a red-light source, and (c) oxygen saturation image SLD-PAM acquired via dual-wavelength spectrum unmixing. Credit: *Advanced Science* (2023). DOI: 10.1002/advs.202302486



"SLD-PAM enables non-invasive imaging of biological tissue with minimal damage to the subjects, offering a powerful and promising tool for anatomical, functional and molecular imaging," said Professor Wang. "We believe that SLD-PAM can help advance the applications of <u>photoacoustic</u> imaging, enable numerous new biomedical applications, and pave a new avenue for clinical translation."

Next, Professor Wang and his research team will test a broader range of small molecules and genetically encoded biomarkers in biological imaging using the SLD-PAM system. They also plan to adopt more types of <u>low-power</u> light sources in a wider spectra to develop wearable or portable microscopy.

**More information:** Yachao Zhang et al, Super-Low-Dose Functional and Molecular Photoacoustic Microscopy, *Advanced Science* (2023). DOI: 10.1002/advs.202302486

## Provided by City University of Hong Kong

Citation: Researchers develop ultra-sensitive photoacoustic microscopy for wide biomedical application potential (2023, August 31) retrieved 9 May 2024 from <u>https://medicalxpress.com/news/2023-08-ultra-sensitive-photoacoustic-microscopy-wide-biomedical.html</u>

This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.