

# New electrode design promises major improvements in wearable electrocardiograms

October 31 2023

---



Illustrative image of ECG device and various applications. Credit: Peter Francis Mathew Elango

Nearly 200 million people around the globe have coronary heart disease,

which accounts for about one in every six deaths, according to the British Heart Foundation. That's why the recent and rapid rise in wearable electronic health-monitoring devices with heart rate-measuring electrocardiograms (ECG) represents a significant step forward. By detecting cardiovascular ailments and helping assess overall cardiac health, wearable ECGs save lives, not to mention exorbitant hospital care costs.

A group of researchers from Australia and India have advanced the positive momentum with a study that presents a novel wearable ECG patch for enhanced point-of-care diagnostics. The article, "Dry electrode geometry optimization for wearable ECG devices," appeared in *Applied Physics Reviews* on Oct. 31, 2023.

"We did a deep dive into the world of wearable ECG devices," said author Peter Francis Mathew Elango. "We focused on a critical aspect, how the design and materials of the electrodes impact their performance, and the findings were quite illuminating."

Typically, Ag/AgCl electrodes, also known as "wet" electrodes, are used in devices to measure ECG signals and include a conductive gel to enhance electrical signaling. But these are known to be uncomfortable, irritating to the skin, and prone to drying out.

Instead, the study focused on the advantages of using active dry electrodes for ECG signaling and ultimately demonstrated an effective and promising prototype.

"Dry electrodes offer some significant advantages," said Elango. "They prioritize user comfort, remain durable over time, and reduce the likelihood of skin irritation. These attributes make them ideal for continuous monitoring, a crucial feature for wearable ECG devices."

The research included considerations of how the performance of dry electrode sensors is affected by the contact area, conductivity, and stability of the electrical contact. The team investigated various tight space geometry-based dry electrodes and different [electrode](#) geometries, studying their various influences on ECG sensing. They also examined how different body parts influence ECG measurements.

After extensive experimentation, they created a compact, lightweight, gel-free hexagonal-shaped ECG patch ideally suited for point-of-care diagnostics. The configuration was then integrated with wireless Bluetooth communication for remote sensing capabilities.

"The potential applications are where it gets really exciting," said Elango. "These portable ECG patches could revolutionize remote and ambulatory health care, and perhaps even preventive medicine. They're light, they can wirelessly transmit data, and they're adept at discerning various states of rest or activity. Whether used in a [clinical setting](#), seamlessly integrated into everyday attire, or worn as [wearable](#) devices, they could redefine how we monitor cardiac health."

**More information:** Peter Francis Mathew Elango et al, Dry electrode geometry optimization for wearable ECG devices, *Applied Physics Reviews* (2023). [DOI: 10.1063/5.0152554](https://doi.org/10.1063/5.0152554).  
[aip.scitation.org/doi/full/10.1063/5.0152554](https://aip.scitation.org/doi/full/10.1063/5.0152554)

Provided by American Institute of Physics

Citation: New electrode design promises major improvements in wearable electrocardiograms (2023, October 31) retrieved 11 May 2024 from <https://medicalxpress.com/news/2023-10-electrode-major-wearable-electrocardiograms.html>

This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.