


Multiple cone-beam scans fall within acceptable ranges

14 March 2013



The radiation dose imparted to patients undergoing multiple intraoperative lumbar single cone-beam computed tomography scans is within the dose range that patients receive during a single fan-beam abdominal scan, according to research published in the March 1 issue of *Spine*.

(HealthDay)—The radiation dose imparted to patients undergoing multiple intraoperative lumbar single cone-beam computed tomography (CT) scans is within the dose range that patients receive during a single fan-beam abdominal CT scan, according to research published in the March 1 issue of *Spine*.

Jeffrey Lange, M.D., of the University of Massachusetts Medical School in Worcester, and colleagues conducted an observational study to estimate the [radiation exposure](#) for patients undergoing cone-beam CT scans during several different thoracolumbar [spinal surgery](#) scenarios.

The researchers found that, for small patients requiring up to six cone-beam CT scans during full-length procedures, the radiation dose ranged from 1 to 31 mSv, which is well within the range of published effective doses received by patients undergoing abdominal CT scans. For large patients, the radiation dose was also within the range, as long as the number of scans did not exceed three.

"Single cone-beam CT scans and most full-length

posterior instrumented spinal procedures using O-arm in standard mode would likely impart a radiation dose within the range of those imparted by a single standard CT scan of the abdomen," the authors write. "Radiation dose increases with patient size, and the [radiation dose](#) received by larger patients as a result of more than three O-arm scans in standard mode may exceed the dose received during standard CT of the abdomen."

This research was funded by [Medtronic](#).

More information: [Abstract](#)
[Full Text \(subscription or payment may be required\)](#)

[Health News](#) Copyright © 2013 [HealthDay](#). All rights reserved.

APA citation: Multiple cone-beam scans fall within acceptable ranges (2013, March 14) retrieved 25 February 2021 from <https://medicalxpress.com/news/2013-03-multiple-cone-beam-scans-fall-ranges.html>

This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.