

Varying radiation exposure in cath lab procedures analyzed

22 October 2013



One author is an employee of Philips Healthcare; Philips provided the DoseAware system used in the study.

More information: [Abstract](#)
[Full Text \(subscription or payment may be required\)](#)

Copyright © 2013 [HealthDay](#). All rights reserved.

(HealthDay)—Endovascular peripheral procedures are associated with higher radiation exposure for catheterization laboratory operators than coronary procedures, according to a study published in the October issue of *JACC: Cardiovascular Interventions*.

Maja Ingwersen, D.V.M., from the University Cardiovascular Center in Hamburg, Germany, and colleagues measured radiation doses of three operators using real-time dosimetry for the body, neck, and hand during 284 procedures in 281 patients over a 14-week period. Mixed models were used to enable comparisons between procedures.

The researchers found that there were independent associations for the type of procedure, the patient's body mass index, and the fluoroscopy time with the operator's radiation exposure. Operators were exposed to a mean effective dose (E) of 2.2 μ Sv per procedure. E was 2.3-fold higher in pelvic procedures (P coronary angiography, upper limb, and below-the-knee procedures).

"Endovascular procedures for pelvic, [upper limb](#), and below-the-knee disease are accompanied with a higher [radiation exposure](#) of the operator than with coronary procedures," the authors write.

APA citation: Varying radiation exposure in cath lab procedures analyzed (2013, October 22) retrieved 27 November 2021 from <https://medicalxpress.com/news/2013-10-varying-exposure-cath-lab-procedures.html>

This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.